



PROPOSAL

Particulate Number Counter

2019-10-16

Instruments for measuring vehicle
exhaust particulate number emissions

Part 1: Metrological and technical requirements

Contents

PART 1 - METROLOGICAL AND TECHNICAL REQUIREMENTS	3
0 Introduction.....	3
1 Scope	3
2 Normative references	3
2.1 ISO Standards	3
2.2 IEC Standards	4
2.3 OIML Publications	4
2.4 UN ECE Publications.....	5
2.4 Other Publications	5
3 Terms and definitions	6
4 Description of the instrument.....	9
5 Metrological requirements	9
5.1 Indication of the measurement result	9
5.2 Measuring range	9
5.3 Resolution of indication	9
5.4 Durable recording of the measurement results	9
5.5 Maximum permissible errors	10
5.6 Influence quantities.....	10
5.7 Disturbances.....	11
5.8 Response time	12
5.9 Warm-up time	12
5.10 Stability with time or drift	12
5.11 Repeatability	12
6 Technical requirements.....	12
6.1 Construction.....	12
6.2 Security of operation.....	14
7 Inscriptions and operating instructions	15
7.1 Inscriptions	15
7.2 Operating instructions	15

Instruments for measuring vehicle exhaust particles

Part 1 - Metrological and technical requirements

0 Introduction

Diesel and petrol engines are more and more equipped with highly effective particulate filters. In order to be able to check the condition of these particulate filters in periodic inspections a simple emission tester and a simple test procedure are needed. Particle counters which measure the volumetric particle concentration are able to detect very low and very high particle emissions and have a linear curve.

- The Proposal will consist of two Parts:
 - Part 1 *Metrological and technical requirements*; and
 - Part 2 *Metrological control and performance tests*.

1 Scope

This Proposal specifies the metrological and technical requirements and tests for digital measuring instruments (hereafter termed “instrument(s)”) that serve to determine the amount of particles per volume of the exhaust gases emanating from combustion engines. The conditions with which such instruments shall comply in order to meet performance requirements are also established.

It is applicable to instruments, intended for the inspection and maintenance of combustion engines. These instruments are used to determine the particle number per volume of defined particle sizes in exhaust gas.

This Proposal does not apply to equipment for on-board diagnostics incorporated in motor vehicles.

2 Normative references

The following normative documents contain provisions that, through reference in this text, constitute provisions of this Proposal.

Subsequent amendments to, or revisions of dated references do not apply. However, parties to agreements based on this Proposal are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies.

Members of ISO, IEC, and the OIML maintain registers of currently valid International Standards. The actual status of the Standards referred to can also be found on the Internet:

- IEC Publications: <http://www.iec.ch>
- ISO Publications: <http://www.iso.org>
- OIML Publications: <http://www.oiml.org> (with free download of PDF files).

2.1 ISO Standards

ISO 3929 (2003), *Road vehicles — Measurement methods for exhaust gas emissions during inspection or maintenance*

ISO 7637-1 (2002) and Amendment 1 (2008), *Road vehicles — Electrical disturbance from conducting and coupling — Part 1: Definitions and general considerations*

ISO 7637-2 (2011) *Road vehicles — electrical disturbance by conducting and coupling – Part 2: Electrical transient conduction along supply lines only*

ISO 7637-3 (2007) *Road vehicles — Electrical disturbance by conducting and coupling — Part 3: Passenger cars and light commercial vehicles with nominal 12 V supply voltage and commercial vehicles with 24 V supply voltage — Electrical transient transmission by capacitive and inductive coupling via lines other than supply lines*

ISO 14912 (2003) with correction 1 (2006), *Gas analysis — conversion of gas mixture composition data*

ISO 15900 (2009) *Determination of particle size distribution -- Differential electrical mobility analysis for aerosol particles*

ISO 16750-2 Ed. 4.0 (2012), *Road vehicles -- Environmental conditions and testing for electrical and electronic equipment -- Part 2: Electrical loads*

ISO 27891 (2015), *Aerosol particle number concentration -- Calibration of condensation particle counters*

2.2 IEC Standards

IEC 60068-2-1 Ed. 6.0 (2007-03), *Environmental testing — Part 2: Test methods — Section 1: Test A: Cold*

IEC 60068-2-2 Ed. 5.0 (2007-07), *Environmental testing — Part 2: Test methods — Section 1: Test B: Dry heat*

IEC 60068-2-30 Ed 3.0 (2005-08), *Environmental testing – Part 2: Test methods – Section 30: Test Db: Damp heat, cyclic (12 + 12 hour cycle)*

IEC 60068-2-78 Ed. 2.0 (2012-10), *Environmental testing — Part 2: Tests methods — Section 78: Test Cab: Damp heat, steady state*

IEC 60068-2-31 Ed. 2.0 (2008-05), *Environmental testing — Part 2: Tests methods — Section 31: Test Ec: Rough handling shocks, primarily for equipment-type specimens*

IEC 60068-2-64 Ed 2.0 (2008-04), *Environmental testing – Part 2: Test methods – Section 64: Test Fh: Vibration, broad-band random and guidance*

IEC 60068-3-1 Ed. 2.0 (2011-08), *Environmental testing — Part 3: Supporting documentation and guidance— Section 1: Cold and dry heat tests*

IEC 60068-3-4 Ed. 1.0 (2001-08), *Environmental testing — Part 3: Supporting documentation and guidance - Section 4: Damp heat tests*

IEC/TR 61000-2-1 Ed. 1.0 (1990-05), *Electromagnetic compatibility (EMC) — Part 2: Environment— Section 1: Description of the environment - Electromagnetic environment for low-frequency conducted disturbances and signaling in public power supply systems*

IEC 61000-4-2 Ed. 2.0 (2008-12), *Basic EMC Publication – Electromagnetic compatibility (EMC) — Part 4: Testing and measurement techniques — Section 2: Electrostatic discharge immunity test*

IEC 61000-4-3 consolidated Ed. 3.2 (2010-04) *Basic EMC Publication – Electromagnetic compatibility (EMC) – Part 4: Testing and measurement techniques – Section 3: Radiated, radio-frequency, electromagnetic field immunity test*

IEC 61000-4-4 Ed. 3.0 (2012-04), *Basic EMC Publication – Electromagnetic compatibility (EMC) – Part 4: Testing and measurement techniques – Section 4: Electrical fast transient/burst immunity test*

IEC 61000-4-5 (2005), Correction 1 on Ed. 2.0 (2009-10), *Basic EMC Publication, Electromagnetic compatibility (EMC) — Part 4: Testing and measurement techniques — Section 5: Surge immunity test*

IEC 61000-4-6 Ed 4.0 (2013-10), *Electromagnetic compatibility (EMC) — Part 4: Testing and measurement techniques — Section 6: Immunity to conducted disturbances, induced by radio-frequency fields*

IEC 61000-4-8 Ed. 2.0 (2009-09). *Basic EMC Publication – Electromagnetic compatibility (EMC) — Part 4: Testing and measurement techniques — Section 8: Power frequency magnetic field immunity test*

2.3 OIML Publications

OIML V 1 (2013) *International vocabulary of terms in legal metrology (VIML)*

OIML V 2-200 (2012) *International Vocabulary of Metrology – Basic and General Concepts and Associated Terms (VIM)*

OIML D 11 (2013) *General requirements for measuring instruments - Environmental conditions*

2.4 UN ECE Publications

UNECE Regulation 83 , *Uniform provisions concerning the approval of vehicles with regard to the emission of pollutants according to engine fuel requirements. Annex 4a, Appendix 5.*

In 2011 a Particle Number measurement was introduced in the European type approval emission legislation for light-duty diesel vehicles. Vehicles have been tested on a chassis dynamometer with a so-called Constant Volume Sampler (CVS) system. The CVS system which operates with highly diluted exhaust gas is equipped with a particulate counter and a sample conditioning system. The sample conditioning system removes the volatile particles. The remaining volumetric solid particle concentration is measured in the particle counter which has a size dependant counting efficiency and a measuring range of around 0 to 25000 particles/cm³.

The type approval emission test consists of a defined driving cycle and the volumetric PN emission of the diluted sample is continuously measured. The final test result is expressed in particles per kilometre.

Due to the very high efficiency of the applied particle filters in vehicles the PN concentration of undiluted exhaust gas at idle speed is around or below the PN concentration of ambient air. DPF leakage will lead to an increased PN emission and a DPF removal mostly yields a PN emission at idle speed in the range of 1 to 15 million particles/cm³.

2.5 Other Publications

Guide to the expression of uncertainty in measurement (GUM), (1995): Joint publication by the BIPM, IEC, IFCC, ISO, IUPAC, IUPAP, and OIML

Verordnung des EJPD über Abgasmessgeräte für Verbrennungsmotoren

3 Terms and definitions

For the purposes of this Proposal, the following terms and definitions apply.

3.1 sampling probe

tube that is introduced into the exhaust tail pipe of a vehicle to take gas samples

3.2 HEPA filter

device that removes particulates from the air

Note: HEPA stand for High Efficiency Particulate Air

3.3 (electrical) mobility diameter (particle size)

size of charged particles equivalent to the electrical mobility in an electrical field

3.4 gas handling system

instrument components, from the sampling probe to the gas sample outlet, through which the exhaust gas sample is conveyed by the pump

3.5 adjustment (of a measuring system)

set of operations carried out on a measuring system so that it provides prescribed indications corresponding to given values of a quantity to be measured (VIML, 0.15)

3.6 semi-automatic adjustment facility

facility allowing the user to initiate an adjustment of the instrument without having the possibility of influencing its magnitude, whether or not the adjustment is automatically required

Note: For those instruments that require the values of the adjustment reference to be entered manually, the facility is considered to be semi-automatic.

3.7 automatic adjustment facility

facility performing the adjustment of the instrument as programmed without the intervention of the user, to initiate the adjustment or its magnitude

3.8 zero-setting facility

facility to set the indication of the instrument to zero

3.9 internal adjustment facility

facility to adjust the instrument to a designated value without the use of an external adjustment reference

3.10 warm-up time

elapsed time between the moment power is applied to an instrument and the moment at which the instrument is capable of complying with the metrological requirements

3.11 (step) response time

duration between the instant when an input quantity value of a measuring instrument or measuring system is subjected to an abrupt change between two specified constant quantity values and the instant when a corresponding indication settles within specified limits around its final steady value (VIM, 4.23)

3.12 reference (quantity) value

quantity value used as a basis for comparison with values of quantities of the same kind (VIM, 5.18)

3.13 (measurement) error

measured quantity value minus a reference quantity value (VIM, 2.16)

3.14 relative error

error of measurement divided by the reference quantity value of the measurand

3.15 fault

difference between the error of indication and the intrinsic error of the instrument (VIML, 5.14)

3.16 significant fault

fault exceeding the applicable fault limit value (VIML, 5.14)

Note: The following faults are considered to be not significant:

- a) fault arising from simultaneous and mutually independent causes in the instrument itself or in its checking facilities;
- b) faults implying the impossibility to perform any measurement;
- c) transitory faults being momentary variations in the indication, which cannot be interpreted, recorded or transmitted as a measurement result; and
- d) faults giving rise to variations in the measurement results that are so large as to be noticed by all those interested in the measurement result.

3.17 influence quantity

quantity that, in a direct measurement, does not affect the quantity that is actually measured, but affects the relation between the indication and the measurement result (VIM, 2.52)

3.18 rated operating conditions

operating condition that must be fulfilled during measurement in order that a measuring instrument or measuring system perform as designed (VIM, 4.9)

3.19 disturbance

influence quantity having a value within the limits specified in this Proposal but outside the rated operating conditions of the instrument (VIML, 5.19)

3.20 reference condition

operating condition prescribed for evaluating the performance of a measuring instrument or measuring system or for comparison of measurement results (VIM, 4.11)

3.21 checking facility

facility that is incorporated in the instrument and that enables significant faults to be detected and acted upon (VIML, 5.07)

Note: “Acted upon” refers to any adequate response by the instrument (luminous signal, acoustic signal, prevention of the measurement process, etc.);

3.22 automatic checking facility

checking facility operating without the intervention of the user

3.23 legally relevant software

any part of the software, including stored parameters, which has an influence on the calculated, displayed, transmitted, or stored measurement result.

3.24 reference PN sample

aerosol of sufficient stability and homogeneity whose composition is properly established for use in various performance tests

3.25 instrument construction types:

a. hand-held instrument

type of instrument that is designed for hand-held use with its standard accessories by one person

b. transportable instrument

type of instrument that rests on a suitable surface during use, which may be designed for hand-held transportation with its standard accessories by one person

3.26 motor vehicle

road vehicle, powered by a built-in engine, which is not rail borne

Note:, normally used for applications such as:

- carrying persons and/or goods;
- towing vehicles used for the carriage of persons and/or goods.

3.27 Abbreviations

AC	Alternating Current
AM	Amplitude Modulation
ASD	Acceleration Spectral Density
DC	Direct Current
EMC	Electro Magnetic Compatibility
e.m.f.	electromotive force
EUT	Equipment Under Test
IEC	International Electrotechnical Commission
ISO	International Organization for Standardization
MPE	Maximum Permissible Error
OIML	International Organization of Legal Metrology
PN	Particle Number

4 Description of the instrument

4.1 Generally, the instrument provides a means for sampling and then measuring the exhaust gases emitted from the tail pipe of a motor vehicle. A pump provides the means for conveying the gas sample through a gas handling system. One or more detection devices, incorporated in the gas handling system, analyze the sample and provide signals related to the number of particles per volume. The detector signals are then electrically processed to display and possibly record the results of a particulate number measurement. With particles is meant solid particles with sizes between 23 and 200 nm.

4.2 The major instrument components are as follows:

- sampling probe introduced in the tail pipe of an operating motor vehicle to collect the exhaust gas sample;
- pump(s) to convey the gases through the instrument;
- device(s) to prevent water condensation from forming in the sampling line and in the instrument;
- filter(s) to remove particles that could cause contamination of various sensitive parts of the instrument;
- ports to introduce ambient air and reference PN sample when required by the technology used;
- device to remove volatile particles of the sample,
- detection devices to measure the particulate number of the gas sample;
- a data system to process the signal including an indicating device to display the results of a measurement; and
- a control facility to initiate and check instrument operations and a semi-automatic or automatic adjustment facility to set instrument operating parameters within prescribed limits.

5 Metrological requirements

5.1 Indication of the measurement result

The particle number per volume shall be expressed as number of particles per cm^3 for particles with specified sizes. The inscriptions for this unit shall be assigned unambiguously to the indication, for example “#/ cm^3 ”.

5.2 Measuring range

The minimum range, that may be subdivided, shall be 5 000 to 5 000 000 particles/ cm^3 .

The exceeding of the range is indicated (visibly) by the instrument.

5.3 Resolution of indication

As indicated in the Scope, this Proposal only refers to digital indicating instruments. Digital figures shall be at least 5 mm high. The least significant figure of the display shall provide a minimum resolution of 1000 particles/ cm^3 .

5.4 Durable recording of the measurement results

The measurement results shall be recorded by a durable means, accompanied by information to identify the particular measurement.

5.5 Maximum permissible error

5.5.1 Maximum permissible error

Type of error	Maximum permissible error *
	[particles/cm ³]
Absolute	25 000
Relative	± 25% of the actual value
* Absolute or relative, whichever is greater.	

5.5.2 Fault limit

The fault limit is 25 000 particles/cm³.

5.6 Influence quantities

5.6.1 Reference conditions

- | | |
|-------------------------|---------------------------------|
| a) ambient temperature | 20 °C ± 2 °C; |
| b) relative humidity | 50 % ± 20 %; |
| c) atmospheric pressure | stable ambient; |
| d) mains voltage | nominal voltage ± 2 %; |
| e) mains frequency | nominal frequency ± 1 %; |
| f) vibration | none / negligible; |
| g) voltage of battery | nominal voltage of the battery. |

5.6.2 Rated operating conditions

- | | |
|------------------------------------|--|
| a) ambient temperature | + 5 °C to + 40 °C ⁽¹⁾ ; |
| b) relative humidity | up to 85 %, no condensation (closed location);
up to 95 % condensing (open location) ⁽²⁾ ; |
| c) atmospheric pressure | 860 hPa to 1 060 hPa ⁽³⁾ ; |
| d) mains voltage | - 15 % to + 10 % of the nominal voltage; |
| e) mains frequency | ± 2 % of the nominal frequency; |
| f) voltage of road vehicle battery | 12 V battery: 9 V to 16 V;
24 V battery: 16 V to 32 V; |
| g) voltage of internal battery | low voltage as specified by the manufacturer, up to the
voltage of a new or fully charged battery of the specified
type. |

- ⁽¹⁾ Unless otherwise specified by the manufacturer, these are the standardized ranges for the ambient temperature. The manufacturer, however, can specify different ranges under the following conditions:
- the lower temperature shall be 5 °C or less;
- the higher temperature shall be either 40 °C or more.

Outside the temperature range the instrument shall not indicate the measured value, instead it shall indicate a warning the temperature is outside the range.

- ⁽²⁾ Closed location: instrument is meant for use inside, open location: instrument is meant for use outside.

- (3) Unless otherwise specified by the manufacturer, this is the standardized range for the atmospheric pressure. The manufacturer can specify an extended range for atmospheric pressure that includes the standardized range.

5.6.3 Influence of particle size

The design of the instrument shall be such that the detection efficiency related to the particle size is the following

Counting efficiency	Particle size [nm]
0,2 – 0,6	23 +/- 5%
0,6 – 1,3	50 +/- 5%
0,7 – 1,3	80 +/- 5%

5.7 Disturbances

Significant faults as defined in 3.21 shall either not occur or shall be detected and acted upon by means of checking facilities in case of the following disturbances:

Mechanical shock	Handheld: 1 fall of 1 m on each bottom edge Transportable: 1 fall of 50 mm on each bottom edge
Vibration	10 Hz to 150 Hz, 1.6 ms ⁻² , 0.05 m ² s ⁻³ , -3 dB/octave
AC mains voltage dips, short interruptions and reductions	0.5 cycles reduction to 0 % 1 cycle reduction to 0 % 10/12 (*) cycles reduction to 40 % 25/30 (*) cycles reduction to 70 % 250/300 (*) cycles reduction to 80 % 250/300 (*) cycles reduction to 0 % (*) For 50 Hz/ 60 Hz respectively
Bursts (transients) on AC mains	Amplitude 2 kV Repetition rate 5 kHz
Bursts (transients) on signal, data and control lines	Amplitude 1 kV Repetition rate 5 kHz
Surges on AC mains power lines	Line to line 1.0 kV Line to ground 2.0 kV
Surges on signal, data and control lines	Line to line 1.0 kV Line to ground 2.0 kV

Electrostatic discharge	6 kV contact discharge 8 kV air discharge
Radiated, radio-frequency, electromagnetic fields	80 (26) MHz up to 6 GHz, 10 V/m
Conducted radio-frequency fields	0.15 up to 80 MHz, 10 V (e.m.f.)
Power frequency magnetic fields	Continuous 100 A/m Short duration 1000 A/m for 1 s
For instruments powered by a road vehicle battery:	
Electrical transient conduction along supply lines	Pulses 2a, 2b, 3a, 3b, test level IV (ISO 7637-2)
Electrical transient conduction via lines other than supply lines	Pulses a and b, test level IV (ISO 7637-3)
Load dump	Test B (ISO 16750-2)

5.8 Response time

For measuring PN concentration, the instrument including the specified gas handling system shall indicate 95 % of the final value (as determined with reference PN samples) within 15 s after changing from ambient air. The instrument may be provided with a logging device to check this requirement.

5.9 Warm-up time

After the warm-up time, the instrument shall meet the metrological requirements stated in this Proposal.

The instrument shall prevent an indication of measured particle number during the warm-up time.

5.10 Stability with time or drift

When used in accordance with the manufacturer's operating instructions, the measurements made by the instrument, under stable environmental conditions and after adjustment using a reference PN sample or the internal adjustment facility, shall remain within the maximum permissible error for at least 12 h without the need for reference PN sample or internal readjustments by the user. If the instrument is equipped with a means for drift compensation, such as automatic zero or automatic internal adjustment, the action of these adjustments shall not produce an indication that can be confused with a measurement of an external gas.

5.11 Repeatability

For 20 consecutive measurements of the same reference PN sample carried out by the same person with the same instrument within relatively short time intervals, the experimental standard deviation of the 20 results shall not be greater than one third of the modulus of the maximum permissible error for the relevant sample.

6 Technical requirements

6.1 Construction

6.1.1 All components of the gas handling system shall be made of corrosion-resistant material. The material of the sampling probe shall withstand the exhaust gas temperature. The materials used shall not influence the composition of the gas sample.

6.1.2 The sampling probe shall be so designed that it can be inserted at least 30 cm into the exhaust tail pipe of the vehicle and be held in place by a retaining device regardless of the depth of insertion.

6.1.3 The instrument shall either contain a device that prevents water condensation from forming in the sampling line and measuring components or a detector that gives an alarm and prevents a measurement result to be indicated.

Note: Examples of devices are: heating of sampling line or dilution with ambient air near the sampling probe.

6.1.4 If an adjustment reference is needed due to the measurement technique simple means to provide such a sample shall be available with the instrument.

6.1.5 The pump conveying the exhaust gas shall be mounted so that its vibrations do not affect the measurements. It is allowed to switch the pump on and off by the user if it is not possible to make a measurement when the pump is switched off.

Note: It is recommended that the gas handling system be flushed automatically with ambient air before the pump is switched off.

6.1.6 The instrument shall be equipped with a device that indicates when the gas flow rate decreases to a level that would cause the detection to exceed either the response time or half the modulus of the maximum permissible error. And, when that limit is reached, the device shall prevent measurements from being carried out.

6.1.7 The gas handling system shall be airtight to such an extent that the influence of dilution with ambient air on the measurement results shall not be more than half the modulus of the maximum permissible error.

The instrument shall not be able to make a measurement if this value is exceeded.

A clean air test procedure with sufficient accuracy (for example HEPA filter with 99,97 % efficiency) to detect this specific maximum leakage shall be provided in the manufacturer's operating instructions.

The instrument shall be equipped with a device that at first use per 24 h performs an automatic zero-setting or zero-setting procedure. This device may be combined with the clean air test procedure.

6.1.8 The instrument may be equipped with an interface permitting coupling to any peripheral device(s) or other instrument(s).

An interface shall not allow the metrological functions of the instrument(s) or their measurement data to be influenced by the peripheral devices, by other interconnected instruments or by disturbances acting on the interface.

Functions that are performed or initiated via an interface shall meet the relevant requirements and conditions.

If the instrument is connected to a data printer or an external data storage device, then the data transmission from the instrument to the printer shall be designed so that the results cannot be falsified.

It shall not be possible to print out a document or store the measuring data in an external device for legal purposes if the instrument checking facility(ies) detect(s) a significant fault or a malfunction.

6.1.9 The instrument has a reporting frequency equal to or greater than a frequency of 1 Hz.

6.1.10 The instrument is designed according to good engineering practice to ensure particle concentration reduction factors (for example of a dilution system) are stable across a vehicle test.

6.1.11 The instrument achieves > 95 percent removal efficiency of 30 nm Tetracontane (C₄₀H₈₂) particles.

6.2 Security of operation

6.2.1 If the detection of one or more of the disturbances listed in 5.7 is achieved by the use of automatic self-checking facilities, then it shall be possible to check the correct functioning of such facilities.

6.2.2 The instrument shall be controlled by an automatic checking facility that shall operate in such a way that, before a measurement can be indicated or printed, all adjustments, and all other checking facility parameters shall be confirmed for proper values or status (i.e. within limits).

6.2.3 Instruments equipped with an automatic adjustment facility or a semi-automatic adjustment facility shall not be able to make a measurement until correct adjustments have been completed.

6.2.4 Instruments equipped with a semi-automatic adjustment facility shall not be able to make a measurement when an adjustment is required.

6.2.5 A means for warning of a required adjustment may be provided for both automatic and semi-automatic adjustment facilities.

6.2.6 Effective sealing devices shall be provided on all parts of the instrument that are not materially protected in another way against operations liable to affect the accuracy or the integrity of the instrument.

This applies in particular to:

- adjustment means;
- software integrity.

6.2.7 The legally relevant software shall be clearly identified. The identification shall be displayed or printed:

- on command or
- during operation or
- at start up for a measuring instrument that can be turned off and on again.

6.2.8 Software shall be protected in such a way that evidence of any intervention (e.g. software updates, parameters changes) shall be available.

6.2.9 The metrological characteristics of an instrument shall not be influenced in any inadmissible way by the connection to it of another device, by any feature of the connected device itself or by any remote device that communicates with the measuring instrument.

6.2.10 A battery-operated instrument shall function correctly with new or fully charged batteries of the specified type and either continue to function correctly or not indicate any values whenever the voltage is below the manufacturer's specified value.

Specific voltage limits for road vehicle batteries are prescribed in 5.6.2.

7 Inscriptions and operating instructions

7.1 Inscriptions

7.1.1 The instrument shall have a permanent, non-transferable, and easily readable label or labels giving the following information:

- a) Manufacturer's trademark/corporate name;
- b) Year of manufacture;
- c) Type approval mark and model number;
- d) Serial number of the instrument;
- e) Details of the electrical power:
 - in case of mains power: the nominal mains voltage, frequency and power required;
 - in case of power by a road vehicle battery: the nominal battery voltage and power required;
 - in case of internal removable battery: the type and nominal voltage of the battery.

7.2 Operating instructions

7.2.1 The manufacturer shall provide written operating instructions for each instrument in the language(s) of the country in which it will be used.

7.2.2 The operating instructions shall include:

- a) The time intervals and the procedures for adjustment and maintenance that shall be followed to comply with the maximum permissible errors (see also 6.2.5);
- b) A description of the clean air test procedure;
- c) If applicable, the zero-setting procedure;
- d) The maximum and minimum storage temperatures;
- e) A statement of the rated operating conditions listed in 5.6.2 and other relevant mechanical and electromagnetic environmental conditions;
- f) If applicable, details about compatibility with ancillary equipment;
- g) The ambient temperature range shall be included in the operating instructions;
- h) Any specific operating conditions, for example a limitation of the length of signal, data, or control lines;
- i) If applicable, the specifications of the battery (see 6.2.10);
- j) A list of error messages with explanation.